

CERTIFICATE OF ANALYSIS

Product Name: Ready-to-use BCA Protein Assay Kit
Cat.No: BH5484
Lot.No: 11AC1005
Production Date: 12/19/2025
Expiry Date: 12/19/2026
Storage Temperature: BSA standard: -20°C; Other components: 4°C

Parameter	Specification	Result
Visual Appearance	Colorless or light yellow, transparent, free of particulates	Conforms
pH	11.25 ± 0.10	11.3
OD of standard NO.1	< 0.150	0.1108
R ²	>0.99	0.9975

Quality authorized by:




Issue date: 12/19/2025

YamayBio Inc.

Address: 6191 Cornerstone Court, Suite 107, San Diego, CA 92121, USA

Website: www.yamaybio.com

Email: info@yamaybio.com